

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
 Data File : VX006012.D
 Acq On : 15 Nov 2018 08:59
 Operator : JC/MD
 Sample : J5925-13ME
 Misc : 13.01g/5ml/100ul/5.00mL/MSVOA_X/MEOH
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6S1ME

Quant Time: Nov 15 10:11:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	118462	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	178382	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	171194	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	92602	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	70645	51.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.74%	
35) Dibromofluoromethane	5.50	113	55437	45.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.38%	
50) Toluene-d8	8.71	98	219313	54.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.74%	
62) 4-Bromofluorobenzene	11.14	95	84776	58.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	116.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
30) Chloroform	5.21	83	3958	1.88	ug/l	97
40) Benzene	6.13	78	6107	1.22	ug/l	93
67) Ethyl Benzene	10.26	91	2227	0.40	ug/l	92
68) m/p-Xylenes	10.36	106	1340	0.64	ug/l	93
69) o-Xylene	10.70	106	1967	1.05	ug/l	89
80) 1,3,5-Trimethylbenzene	11.51	105	1583	0.35	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	7765	1.43	ug/l	100
95) Naphthalene	13.83	128	147391	22.78	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111418\
Data File : VX006012.D
Acq On : 15 Nov 2018 08:59
Operator : JC/MD
Sample : J5925-13ME
Misc : 13.01g/5ml/100ul/5.00mL/MSVOA_X/MEOH
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6S1ME

Quant Time: Nov 15 10:11:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

