

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111419\
 Data File : VX013412.D
 Acq On : 14 Nov 2019 11:09
 Operator : JC/SP
 Sample : K5804-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB111219

Quant Time: Nov 15 02:15:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.00	128	120056	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	655192	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	569824	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.05	65	257626	29.06	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	96.87%
60) 4-Bromofluorobenzene	11.13	95	250995	27.41	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	91.37%
63) Toluene-d8	8.71	98	777856	30.19	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.63%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111419\
Data File : VX013412.D
Acq On : 14 Nov 2019 11:09
Operator : JC/SP
Sample : K5804-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB111219

Quant Time: Nov 15 02:15:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 12 01:16:13 2019
Response via : Initial Calibration

