

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111419\
 Data File : VX013426.D
 Acq On : 14 Nov 2019 16:46
 Operator : JC/SP
 Sample : K5828-02DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GE-MW-57A-2DL

Quant Time: Nov 15 04:51:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	118366	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	644339	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	574613	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.05	65	257109	29.42	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	98.07%
60) 4-Bromofluorobenzene	11.14	95	250321	27.10	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.33%
63) Toluene-d8	8.71	98	769603	29.62	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.73%

Target Compounds

					Ovalue
24) Methyl tert-Butyl Ether	3.19	73	69113	2.857 ug/l	98
37) Trichloroethene	7.21	130	398379	49.815 ug/l	98
61) Tetrachloroethene	9.34	164	6821	0.963 ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111419\
Data File : VX013426.D
Acq On : 14 Nov 2019 16:46
Operator : JC/SP
Sample : K5828-02DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GE-MW-57A-2DL

Quant Time: Nov 15 04:51:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 12 01:16:13 2019
Response via : Initial Calibration

