

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111419\
 Data File : VX013413.D
 Acq On : 14 Nov 2019 11:32
 Operator : JC/SP
 Sample : VX1114WBS01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Manual Integrations
 APPROVED

MMDadoda
 11/15/2019 9:21:03 AM

Quant Time: Nov 15 02:19:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	112692	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	593135	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	531368	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.05	65	241971	29.08	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	96.93%
60) 4-Bromofluorobenzene	11.14	95	255673	29.94	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	99.80%
63) Toluene-d8	8.71	98	720808	30.00	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.00%

Target Compounds

					Ovalue
2) Dichlorodifluoromethane	1.19	85	135543	16.053 ug/l	98
3) Chloromethane	1.32	50	166422	18.085 ug/l	100
4) Vinyl Chloride	1.40	62	176359	16.866 ug/l	99
5) Bromomethane	1.63	94	105089	16.718 ug/l	97
6) Chloroethane	1.72	64	101743	17.701 ug/l	100
7) Trichlorofluoromethane	1.92	101	185130	16.446 ug/l	98
8) Diethyl Ether	2.18	74	100789	19.609 ug/l	99
9) 1,1,2-Trichlorotrifluoroet	2.37	101	110451	16.308 ug/l	100
10) 1,1-Dichloroethene	2.36	96	115192	16.551 ug/l	97
11) Methyl Iodide	2.50	142	145601	16.943 ug/l	99
12) Methyl Acetate	2.76	43	237685	19.489 ug/l	98
13) Acrolein	2.28	56	118691	91.621 ug/l	98
14) Acrylonitrile	3.13	53	427568	95.341 ug/l	98
15) Acetone	2.43	58	113117	83.281 ug/l	95
16) Carbon Disulfide	2.56	76	307864	15.859 ug/l	99
17) Allyl chloride	2.72	41	224172	17.490 ug/l	96
18) Methylene Chloride	2.85	84	148086	18.753 ug/l	99
19) trans-1,2-Dichloroethene	3.15	96	130051	17.484 ug/l	97
20) Diisopropyl ether	3.84	45	468368	18.994 ug/l	91
21) 1,1-Dichloroethane	3.69	63	243774	18.151 ug/l	98
22) cis-1,2-Dichloroethene	4.58	96	156231	18.423 ug/l	96
23) tert-Butyl Alcohol	3.03	59	199880	92.613 ug/l	# 100
24) Methyl tert-Butyl Ether	3.19	73	438580	19.040 ug/l	100
25) Chloroform	5.20	83	240981	18.331 ug/l	98
26) Cyclohexane	5.57	56	198426	16.269 ug/l	# 98
29) 1,1-Dichloropropene	5.79	75	167077	18.442 ug/l	99
30) 2-Butanone	4.66	43	613439	102.586 ug/l	99
31) 2,2-Dichloropropane	4.58	77	175956	17.934 ug/l	99
32) 1,1,1-Trichloroethane	5.48	97	188504	18.720 ug/l	98
33) Carbon Tetrachloride	5.78	117	154315	17.643 ug/l	99
34) Benzene	6.14	78	553379	20.138 ug/l	98
35) Methacrylonitrile	5.03	41	128610	21.537 ug/l	97
36) 1,2-Dichloroethane	6.18	62	205777	21.394 ug/l	100
37) Trichloroethene	7.21	130	141071	19.163 ug/l	95
38) Methylcyclohexane	7.46	83	194988	17.553 ug/l	98
39) 1,2-Dichloropropane	7.51	63	144090	20.280 ug/l	98
40) Dibromomethane	7.65	93	96374	20.381 ug/l	97

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111419\
 Data File : VX013413.D
 Acq On : 14 Nov 2019 11:32
 Operator : JC/SP
 Sample : VX1114WBS01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Manual Integrations
 APPROVED

MMDadoda
 11/15/2019 9:21:03 AM

Quant Time: Nov 15 02:19:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Nov 12 01:16:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Bromodichloromethane	7.89	83	186637	20.219	ug/l	98
42) Vinyl Acetate	3.80	43	1981045	105.226	ug/l	100
43) Ethyl Acetate	4.82	43	232289	21.299	ug/l	99
44) Isopropyl Acetate	6.43	43	358598	20.455	ug/l	99
45) 1,4-Dioxane	7.73	88	78168	403.857	ug/l	98
46) Methyl methacrylate	7.76	41	181574	20.821	ug/l	98
47) n-amyl Acetate	10.89	43	303190	19.716	ug/l	98
48) t-1,3-Dichloropropene	9.04	75	201007	19.420	ug/l	98
49) cis-1,3-Dichloropropene	8.43	75	223600	19.715	ug/l	99
50) 1,1,2-Trichloroethane	9.21	97	150227	21.470	ug/l	97
51) Ethyl methacrylate	9.17	69	236079	20.492	ug/l	97
52) 1,3-Dichloropropane	9.37	76	254020	21.236	ug/l	99
53) Dibromochloromethane	9.57	129	151326	20.406	ug/l	96
54) 1,2-Dibromoethane	9.67	107	154628	20.667	ug/l	97
55) 2-Chloroethyl vinyl ether	8.31	63	494184	100.392	ug/l	100
56) Bromoform	10.85	173	115315	19.539	ug/l	99
58) 4-Methyl-2-Pentanone	8.64	43	1186674	106.904	ug/l	100
59) 2-Hexanone	9.48	43	905508	103.098	ug/l	100
61) Tetrachloroethene	9.33	164	124256	18.963	ug/l	98
62) Toluene	8.78	91	585359	19.777	ug/l	100
64) Chlorobenzene	10.14	112	368517	19.913	ug/l	99
65) 1,1,1,2-Tetrachloroethane	10.21	131	137698	19.877	ug/l	99
66) Ethyl Benzene	10.25	91	617946	19.004	ug/l	99
67) m/p-Xylenes	10.35	106	473341	37.737	ug/l	99
68) o-Xylene	10.70	106	234512	19.373	ug/l	100
69) Styrene	10.71	104	399384	19.322	ug/l	99
70) Isopropylbenzene	11.01	105	601091	18.845	ug/l	99
71) 1,1,2,2-Tetrachloroethane	11.26	83	242210	21.208	ug/l	98
72) 1,2,3-Trichloropropane	11.29	75	192296m	19.941	ug/l	
73) Bromobenzene	11.25	156	167609	19.926	ug/l	97
74) n-propylbenzene	11.35	91	664367	18.328	ug/l	99
75) 2-Chlorotoluene	11.42	91	413515	19.183	ug/l	100
76) 1,3,5-Trimethylbenzene	11.50	105	496286	18.511	ug/l	98
77) t-1,4-Dichloro-2-butene	11.07	75	73509	18.852	ug/l	98
78) 4-Chlorotoluene	11.51	91	466316	18.881	ug/l	99
79) tert-butylbenzene	11.76	119	487608	18.026	ug/l	99
80) 1,2,4-Trimethylbenzene	11.80	105	496508	18.583	ug/l	99
81) sec-Butylbenzene	11.94	105	549890	17.817	ug/l	99
82) p-Isopropyltoluene	12.06	119	500556	17.603	ug/l	99
83) 1,3-Dichlorobenzene	12.02	146	275666	18.640	ug/l	99
84) 1,4-Dichlorobenzene	12.09	146	274423	18.694	ug/l	99
85) n-Butylbenzene	12.39	91	392638	15.998	ug/l	99
86) Hexachloroethane	12.59	117	89728	17.195	ug/l	98
87) 1,2-Dichlorobenzene	12.39	146	285214	19.287	ug/l	98
88) 1,2-Dibromo-3-Chloropropan	12.99	75	51361	20.109	ug/l	95
89) 1,2,4-Trichlorobenzene	13.64	180	169719	16.645	ug/l	98
90) Hexachlorobutadiene	13.78	225	78870	16.142	ug/l	97
91) Naphthalene	13.83	128	580143	17.970	ug/l	99
92) 1,2,3-Trichlorobenzene	14.01	180	181420	17.606	ug/l	99

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111419\
Data File : VX013413.D
Acq On : 14 Nov 2019 11:32
Operator : JC/SP
Sample : VX1114WBS01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01


Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM

Quant Time: Nov 15 02:19:37 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X111119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Nov 12 01:16:13 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1





Bromochloromethane
Concen: 30.000 ug/l
RT: 5.01 min Scan# 676
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Resp	Lower	Upper
128	100	112692		
49	159.2	0.0	388.3	
130	130.1	0.0	318.5	

Manual Integrations
APPROVED




MMDadoda
11/15/2019 9:21:03 AM




#2

Dichlorodifluoromethane
Concen: 16.053 ug/l
RT: 1.19 min Scan# 50
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Resp	Lower	Upper
85	100	135543		
87	31.7	16.6	49.7	





#3

Chloromethane

Concen: 18.085 ug/l

RT: 1.32 min Scan# 71

Delta R.T. 0.01 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X

ClientSampleId :

VX1114WBS01

Tot Ion: 50 Resp: 166422

Ion Ratio Lower Upper





50 100

52 33.5 26.6 40.0

Manual Integrations
APPROVED

MMDadoda

11/15/2019 9:21:03 AM



#4

Vinyl Chloride

Concen: 16.866 ug/l

RT: 1.40 min Scan# 84

Delta R.T. 0.00 min

Lab File: VX013413.D




Acq: 14 Nov 2019 11:32


Tgt Ion: 62 Resp: 176359

Ion Ratio Lower Upper

62 100

64 31.5 25.4 38.2





#5

Bromomethane

Concen: 16.718 ug/l

RT: 1.63 min Scan# 122

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X

ClientSampleId :

VX1114WBS01

Tot Ion: 94 Resp: 105089

Ion Ratio Lower Upper

94 100





96 91.8 75.7 113.5

Manual Integrations

APPROVED

MMDadoda

11/15/2019 9:21:03 AM



#6

Chloroethane

Concen: 17.701 ug/l

RT: 1.72 min Scan# 136

Delta R.T. 0.00 min

Lab File: VX013413.D




Acq: 14 Nov 2019 11:32


Tgt Ion: 64 Resp: 101743

Ion Ratio Lower Upper

64 100

66 31.9 25.5 38.3





#7





Trichlorofluoromethane
Concen: 16.446 ug/l
RT: 1.92 min Scan# 170
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Resp	Lower	Upper
101	100	185130		
103	62.8	51.2	76.8	

Manual Integrations
APPROVED




MMDadoda
11/15/2019 9:21:03 AM




#8

Diethyl Ether
Concen: 19.609 ug/l
RT: 2.18 min Scan# 212
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Resp	Lower	Upper
74	100	100789		
45	96.9	19.5	175.5	





#9

1,1,2-Trichlorotrifluoroethane

Concen: 16.308 ug/l

RT: 2.37 min Scan# 244

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X





ClientSampleId :

VX1114WBS01

Tot Ion:	101	Resp:	110451
Ion Ratio	Lower	Upper	
101	100		
85	41.7	0.0	83.2
151	82.3	0.0	163.8

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#10

1,1-Dichloroethene

Concen: 16.551 ug/l




RT: 2.36 min Scan# 242


Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Tgt Ion:	96	Resp:	115192
Ion Ratio	Lower	Upper	
96	100		
61	162.5	128.9	193.3
98	68.2	49.7	74.5









#11
Methyl Iodide
Concen: 16.943 ug/l
RT: 2.50 min Scan# 265
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
142	100		
127	38.1	19.1	57.5
141	15.1	7.0	21.0




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#12
Methyl Acetate
Concen: 19.489 ug/l
RT: 2.76 min Scan# 307
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
43	100		
74	23.3	18.0	27.0





#13

Acrolein

Concen: 91.621 ug/l

RT: 2.28 min Scan# 229

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X

ClientSampleId :

VX1114WBS01

Tot Ion: 56 Resp: 118691

Ion Ratio Lower Upper

56 100





55 67.9 55.8 83.8

Manual Integrations

APPROVED

MMDadoda

11/15/2019 9:21:03 AM



#14

Acrylonitrile

Concen: 95.341 ug/l

RT: 3.13 min Scan# 368

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32




Tgt Ion: 53 Resp: 427568


Ion Ratio Lower Upper

53 100

52 83.8 41.0 123.0

51 35.7 17.3 51.9









#15
Acetone
Concen: 83.281 ug/l
RT: 2.43 min Scan# 253
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
58	100		
43	316.7	261.4	392.2




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#16
Carbon Disulfide
Concen: 15.859 ug/l
RT: 2.56 min Scan# 275
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
76	100		
78	8.8	7.4	11.0









#17
Allvl chloride
Concen: 17.490 ug/l
RT: 2.72 min Scan# 300
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
41	100		
39	69.6	36.1	108.3
76	33.6	18.1	54.4




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#18
Methylene Chloride
Concen: 18.753 ug/l
RT: 2.85 min Scan# 322
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
84	100		
49	121.9	96.8	145.2
51	37.6	30.0	45.0
86	62.7	51.7	77.5





#19

trans-1,2-Dichloroethene

Concen: 17.484 ug/l

RT: 3.15 min Scan# 372

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X





ClientSampleId :

VX1114WBS01

Tot Ion:	96	Resp:	130051
Ion Ratio	Lower	Upper	
96	100		
61	142.2	109.9	164.9
98	63.8	49.7	74.5

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#20

Diisopropyl ether

Concen: 18.994 ug/l




RT: 3.84 min Scan# 485


Delta R.T. 0.01 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Tgt Ion:	45	Resp:	468368
Ion Ratio	Lower	Upper	
45	100		
43	73.4	67.7	101.5
87	27.5	21.5	32.3
59	11.7	8.5	12.7





#21

1,1-Dichloroethane

Concen: 18.151 ug/l

RT: 3.69 min Scan# 459

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X





ClientSampleId :

VX1114WBS01

Tot Ion	Ratio	Lower	Upper
63	100		
98	7.0	4.0	11.9
100	5.2	2.5	7.5

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#22

cis-1,2-Dichloroethene

Concen: 18.423 ug/l




RT: 4.58 min Scan# 606


Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
96	100		
61	140.8	117.0	175.6
98	62.6	51.1	76.7





#23





tert-Butyl Alcohol
Concen: 92.613 ug/l
RT: 3.03 min Scan# 351
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
59	100		
52	0.2	0.1	0.1#

Manual Integrations
APPROVED




MMDadoda
11/15/2019 9:21:03 AM




#24

Methyl tert-Butyl Ether
Concen: 19.040 ug/l
RT: 3.19 min Scan# 377
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
73	100		
43	21.5	17.0	25.6









#25
Chloroform
Concen: 18.331 ug/l
RT: 5.20 min Scan# 708
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Resp	Lower	Upper
83	100	240981		
85	66.9	52.4	78.6	




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#26
Cyclohexane
Concen: 16.269 ug/l
RT: 5.57 min Scan# 768
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Resp	Lower	Upper
56	100	198426		
89	0.0	0.0	0.0#	
84	86.7	70.5	105.7	









#27
 1,2-Dichloroethane-d4
 Concen: 29.078 ug/l
 RT: 6.05 min Scan# 847
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Tot Ion	Ratio	Lower	Upper
65	100		
67	53.4	41.0	61.6




Manual Integrations
 APPROVED


MMDadoda
 11/15/2019 9:21:03 AM



#28
 1,4-Difluorobenzene
 Concen: 30.000 ug/l
 RT: 6.85 min Scan# 978
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
114	100		
63	19.5	16.1	24.1
88	15.5	12.3	18.5









#29
 1,1-Dichloropropene
 Concen: 18.442 ug/l
 RT: 5.79 min Scan# 804
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Tot Ion	Ratio	Resp Lower	Upper
75	100		
110	36.7	0.0	74.6
77	30.8	0.0	62.6




Manual Integrations
 APPROVED


MMDadoda
 11/15/2019 9:21:03 AM



#30
 2-Butanone
 Concen: 102.586 ug/l
 RT: 4.66 min Scan# 619
 Delta R.T. 0.01 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Resp Lower	Upper
43	100		
57	7.7	6.1	9.1
72	26.2	20.5	30.7









#31
2,2-Dichloropropane
Concen: 17.934 ug/l
RT: 4.58 min Scan# 605
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
77	100		
97	24.3	19.2	28.8




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#32
1,1,1-Trichloroethane
Concen: 18.720 ug/l
RT: 5.48 min Scan# 754
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
97	100		
99	63.7	52.5	78.7
61	46.7	36.1	54.1





#33

Carbon Tetrachloride

Concen: 17.643 ug/l

RT: 5.78 min Scan# 802

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X





ClientSampleId :

VX1114WBS01

Tot Ion:	117	Resp:	154315
Ion Ratio	Lower	Upper	
117	100		
119	94.0	75.0	112.6
121	29.2	24.9	37.3

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#34

Benzene

Concen: 20.138 ug/l




RT: 6.14 min Scan# 861







Delta R.T. 0.01 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32


Tgt Ion:	78	Resp:	553379
Ion Ratio	Lower	Upper	
78	100		
77	24.5	19.3	28.9
51	16.7	12.2	18.4





#35
Methacrylonitrile
Concen: 21.537 ug/l
RT: 5.03 min Scan# 679
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tot Ion	Ion	Ratio	Resp	Lower	Upper
41	100		128610		
39	51.2	25.9	77.8		
67	65.2	35.1	105.5		
52	25.0	12.5	37.5		




Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01


Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM

#36
1,2-Dichloroethane
Concen: 21.394 ug/l
RT: 6.18 min Scan# 869
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
62	100		205777		
98	10.2	8.2	12.2		









#37
Trichloroethene
Concen: 19.163 ug/l
RT: 7.21 min Scan# 1037
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
130	100		
95	94.0	71.3	106.9




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#38
Methylcyclohexane
Concen: 17.553 ug/l
RT: 7.46 min Scan# 1078
Delta R.T. 0.01 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
83	100		
55	78.2	37.9	113.7
98	50.4	25.1	75.3









#39
 1,2-Dichloropropane
 Concen: 20.280 ug/l
 RT: 7.51 min Scan# 1086
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Tot Ion	Ratio	Lower	Upper
63	100		
65	30.8	23.8	35.8




Manual Integrations
 APPROVED


MMDadoda
 11/15/2019 9:21:03 AM



#40
 Dibromomethane
 Concen: 20.381 ug/l
 RT: 7.65 min Scan# 1110
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
93	100		
95	83.1	0.0	160.8
174	117.9	0.0	228.8





#41

Bromodichloromethane

Concen: 20.219 ug/l

RT: 7.89 min Scan# 1149

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X





ClientSampleId :

VX1114WBS01

Tot Ion:	83	Resp:	186637
Ion Ratio	Lower	Upper	
83	100		
85	62.8	49.3	73.9
127	7.8	6.6	10.0

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#42

Vinyl Acetate

Concen: 105.226 ug/l




RT: 3.80 min Scan# 478


Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Tgt Ion:	43	Resp:	1981045
Ion Ratio	Lower	Upper	
43	100		
86	9.6	7.6	11.4









#43
Ethyl Acetate
Concen: 21.299 ug/l
RT: 4.82 min Scan# 645
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	43	45	Ratio	100	14.0	Resp	232289	Lower	10.9	Upper	16.3
Ion	43	45	Ratio	100	14.0	Resp	232289	Lower	10.9	Upper	16.3




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#44
Isopropyl Acetate
Concen: 20.455 ug/l
RT: 6.43 min Scan# 910
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	43	61	Ratio	100	20.2	Resp	358598	Lower	15.8	Upper	23.8
Ion	43	61	Ratio	100	20.2	Resp	358598	Lower	15.8	Upper	23.8









#45
1,4-Dioxane
Concen: 403.857 ug/l
RT: 7.73 min Scan# 1123
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
88	100		
43	33.7	26.1	39.1
58	71.7	55.8	83.8




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#46
Methyl methacrylate
Concen: 20.821 ug/l
RT: 7.76 min Scan# 1128
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
41	100		
69	80.9	41.4	124.4
39	54.2	27.9	83.6









#47
n-amvl Acetate
Concen: 19.716 ug/l
RT: 10.89 min Scan# 1641
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion: 43 Resp: 303190
Ion Ratio Lower Upper
43 100
55 24.6 19.0 28.4




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#48
t-1,3-Dichloropropene
Concen: 19.420 ug/l
RT: 9.04 min Scan# 1337
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion: 75 Resp: 201007
Ion Ratio Lower Upper
75 100
77 32.5 25.3 37.9









#49
 cis-1,3-Dichloropropene
 Concen: 19.715 ug/l
 RT: 8.43 min Scan# 1237
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Tot Ion	Ratio	Lower	Upper
75	100		
77	33.0	25.8	38.8
39	51.6	42.2	63.4




Manual Integrations
 APPROVED


MMDadoda
 11/15/2019 9:21:03 AM



#50
 1,1,2-Trichloroethane
 Concen: 21.470 ug/l
 RT: 9.21 min Scan# 1365
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
97	100		
83	84.1	69.6	104.4
85	55.9	45.5	68.3
99	65.9	50.0	75.0





#51

Ethyl methacrylate

Concen: 20.492 ug/l

RT: 9.17 min Scan# 1359

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X





ClientSampleId :

VX1114WBS01

Tot Ion:	69	Resp:	236079
Ion Ratio	Lower	Upper	
69	100		
41	69.9	33.9	101.6
39	34.9	18.8	56.4

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#52

1,3-Dichloropropane

Concen: 21.236 ug/l




RT: 9.37 min Scan# 1391


Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Tgt Ion:	76	Resp:	254020
Ion Ratio	Lower	Upper	
76	100		
78	32.2	0.0	63.6





#53

Dibromochloromethane

Concen: 20.406 ug/l

RT: 9.57 min Scan# 1425

Delta R.T. -0.01 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X

ClientSampleId :

VX1114WBS01

Tot Ion:129 Resp: 151326

Ion Ratio Lower Upper

129 100


127 75.8 39.6 119.0

Manual Integrations

APPROVED

MMDadoda

11/15/2019 9:21:03 AM



Abundance Ion 129.00 (128.70 to 129.70): V

Ion 127.00 (126.70 to 127.70): V

120000

100000

80000

60000



40000

20000

0

9.50 9.55 9.60

Time-->



#54

1,2-Dibromoethane

Concen: 20.667 ug/l

RT: 9.67 min Scan# 1440

Delta R.T. 0.00 min

Lab File: VX013413.D


Acq: 14 Nov 2019 11:32

Tgt Ion:107 Resp: 154628

Ion Ratio Lower Upper

107 100

109 95.3 73.9 110.9



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V

120000

100000

80000

60000


40000


20000

0

9.60 9.65 9.70

Time-->





#55

2-Chloroethyl vinyl ether

Concen: 100.392 ug/l

RT: 8.31 min Scan# 1217

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X





ClientSampleId :

VX1114WBS01

Tot Ion	Ratio	Lower	Upper
63	100		
65	32.6	26.0	39.0
106	29.3	23.5	35.3

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#56

Bromoform

Concen: 19.539 ug/l




RT: 10.85 min Scan# 1635


Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
173	100		
175	48.5	39.4	59.0
254	10.6	8.6	12.8









#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
117	100		
82	55.0	43.7	65.5
119	32.7	25.5	38.3




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#58
4-Methyl-2-Pentanone
Concen: 106.904 ug/l
RT: 8.64 min Scan# 1271
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
43	100		
58	39.3	31.6	47.4
85	17.9	14.2	21.2









#59
2-Hexanone
Concen: 103.098 ug/l
RT: 9.48 min Scan# 1410
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
43	100		
58	54.7	43.8	65.6
57	18.8	15.1	22.7




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#60
4-Bromofluorobenzene
Concen: 29.936 ug/l
RT: 11.14 min Scan# 1681
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
95	100		
174	88.6	70.2	105.2
176	88.1	68.3	102.5





#61

Tetrachloroethene

Concen: 18.963 ug/l

RT: 9.33 min Scan# 1385

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X




ClientSampleId :

VX1114WBS01

Tot Ion	Ratio	Lower	Upper
164	100		
166	131.1	104.9	157.3
129	84.1	72.3	108.5
131	85.9	69.0	103.6

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#62

Toluene

Concen: 19.777 ug/l




RT: 8.78 min Scan# 1295


Delta R.T. 0.01 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
91	100		
92	58.2	46.6	69.8









#63
Toluene-d8
Concen: 29.998 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion: 98 Resp: 720808
Ion Ratio Lower Upper
98 100
100 66.0 52.5 78.7




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#64
Chlorobenzene
Concen: 19.913 ug/l
RT: 10.14 min Scan# 1517
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion: 112 Resp: 368517
Ion Ratio Lower Upper
112 100
114 31.6 25.7 38.5









#65
 1,1,1,2-Tetrachloroethane
 Concen: 19.877 ug/l
 RT: 10.21 min Scan# 1530
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Tot Ion	Ratio	Resp	Lower	Upper
131	100	137698		
133	95.1	0.0	192.2	
119	68.2	0.0	134.4	




Manual Integrations
 APPROVED


MMDadoda
 11/15/2019 9:21:03 AM



#66
 Ethyl Benzene
 Concen: 19.004 ug/l
 RT: 10.25 min Scan# 1535
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Resp	Lower	Upper
91	100	617946		
106	30.9	25.0	37.4	









#67
m/p-Xvlenes
Concen: 37.737 ug/l
RT: 10.35 min Scan# 1553
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion:106 Resp: 473341
Ion Ratio Lower Upper
106 100
91 199.4 158.4 237.6




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#68
o-Xylene
Concen: 19.373 ug/l
RT: 10.70 min Scan# 1609
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion:106 Resp: 234512
Ion Ratio Lower Upper
106 100
91 213.5 106.4 319.2









#69
Stvrene
Concen: 19.322 ug/l
RT: 10.71 min Scan# 1611
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
104	100		
78	49.2	38.2	57.4
103	53.9	43.2	64.8




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#70
Isopropylbenzene
Concen: 18.845 ug/l
RT: 11.01 min Scan# 1661
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
105	100		
120	27.3	18.8	34.8









#71
 1,1,2,2-Tetrachloroethane
 Concen: 21.208 ug/l
 RT: 11.26 min Scan# 1702
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Tot Ion	Ratio	Lower	Upper
83	100		
131	10.0	4.7	14.1
85	63.1	32.6	97.7




Manual Integrations
 APPROVED


MMDadoda
 11/15/2019 9:21:03 AM



#72
 1,2,3-Trichloropropane
 Concen: 19.941 ug/l m
 RT: 11.29 min Scan# 1707
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
75	100		
77	40.6	0.0	86.2





#73

Bromobenzene

Concen: 19.926 ug/l

RT: 11.25 min Scan# 1700

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X





ClientSampleId :

VX1114WBS01

Tot Ion:	156	Resp:	167609
Ion Ratio	Lower	Upper	
156	100		
77	149.9	0.0	308.8
158	97.3	0.0	197.0

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#74

n-propylbenzene

Concen: 18.328 ug/l




RT: 11.35 min Scan# 1717


Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Tgt Ion:	91	Resp:	664367
Ion Ratio	Lower	Upper	
91	100		
120	23.4	0.0	46.2





#75

2-Chlorotoluene

Concen: 19.183 ug/l

RT: 11.42 min Scan# 1727

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X

ClientSampleId :

VX1114WBS01

Tot Ion: 91 Resp: 413515


Ion Ratio Lower Upper

91 100

126 34.4 0.0 69.0

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 126.00 (125.70 to 126.70): V

600000



400000

200000

0

11.35 11.40 11.45

Time-->



#76

1,3,5-Trimethylbenzene

Concen: 18.511 ug/l

RT: 11.50 min Scan# 1741

Delta R.T. 0.00 min

Lab File: VX013413.D


Acq: 14 Nov 2019 11:32

Tgt Ion: 105 Resp: 496286

Ion Ratio Lower Upper

105 100

120 51.1 0.0 100.0



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

400000

300000


200000


100000

0

11.45 11.50 11.55

Time-->





#77

t-1,4-Dichloro-2-butene

Concen: 18.852 ug/l

RT: 11.07 min Scan# 1670

Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X

ClientSampleId :

VX1114WBS01

Tot Ion: 75 Resp: 73509

Ion Ratio Lower Upper

75 100





53 99.2 48.8 146.3

89 43.0 22.3 66.8

Manual Integrations
APPROVED

MMDadoda

11/15/2019 9:21:03 AM



#78

4-Chlorotoluene

Concen: 18.881 ug/l

RT: 11.51 min Scan# 1742

Delta R.T. 0.00 min

Lab File: VX013413.D




Acq: 14 Nov 2019 11:32


Tgt Ion: 91 Resp: 466316

Ion Ratio Lower Upper

91 100

126 31.2 0.0 61.0





#79

tert-butylbenzene

Concen: 18.026 ug/l

RT: 11.76 min Scan# 1784

Delta R.T. -0.01 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Instrument :

MSVOA_X





ClientSampleId :

VX1114WBS01

Tot Ion:	119	Resp:	487608
Ion Ratio	Lower	Upper	
119	100		
91	55.4	0.0	109.6
134	24.2	0.0	47.8

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#80

1,2,4-Trimethylbenzene

Concen: 18.583 ug/l




RT: 11.80 min Scan# 1790


Delta R.T. 0.00 min

Lab File: VX013413.D

Acq: 14 Nov 2019 11:32

Tgt Ion:	105	Resp:	496508
Ion Ratio	Lower	Upper	
105	100		
120	46.5	0.0	94.0









#81
 sec-Butylbenzene
 Concen: 17.817 ug/l
 RT: 11.94 min Scan# 1813
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Tot Ion	Ratio	Lower	Upper
105	100		
134	20.4	0.0	41.2




Manual Integrations
 APPROVED


MMDadoda
 11/15/2019 9:21:03 AM



#82
 p-Isopropyltoluene
 Concen: 17.603 ug/l
 RT: 12.06 min Scan# 1833
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
119	100		
134	26.6	0.0	51.6
91	22.1	0.0	44.2









#83
 1,3-Dichlorobenzene
 Concen: 18.640 ug/l
 RT: 12.02 min Scan# 1826
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Tot Ion	Ratio	Lower	Upper
146	100		
111	38.3	18.8	56.3
148	63.6	32.3	96.9




Manual Integrations
 APPROVED


MMDadoda
 11/15/2019 9:21:03 AM



#84
 1,4-Dichlorobenzene
 Concen: 18.694 ug/l
 RT: 12.09 min Scan# 1838
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
146	100		
111	36.2	18.6	55.6
148	63.8	32.5	97.4









#85
n-Butylbenzene
Concen: 15.998 ug/l
RT: 12.39 min Scan# 1886
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Resp	Lower	Upper
91	100	392638		
92	55.0	0.0	109.0	
134	27.4	0.0	53.0	




Manual Integrations
APPROVED


MMDadoda
11/15/2019 9:21:03 AM



#86
Hexachloroethane
Concen: 17.195 ug/l
RT: 12.59 min Scan# 1920
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Resp	Lower	Upper
117	100	89728		
201	86.7	44.5	133.3	









#87
 1,2-Dichlorobenzene
 Concen: 19.287 ug/l
 RT: 12.39 min Scan# 1886
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBS01

Tot Ion	Ratio	Lower	Upper
146	100		
111	40.0	19.6	58.7
148	65.2	31.6	94.7




Manual Integrations
 APPROVED


MMDadoda
 11/15/2019 9:21:03 AM



#88
 1,2-Dibromo-3-Chloropropane
 Concen: 20.109 ug/l
 RT: 12.99 min Scan# 1986
 Delta R.T. 0.00 min
 Lab File: VX013413.D
 Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
75	100		
155	92.0	76.9	115.3
157	119.7	101.2	151.8









#89
1,2,4-Trichlorobenzene
Concen: 16.645 ug/l
RT: 13.64 min Scan# 2092
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Resp	Lower	Upper
180	100	169719		
182	96.1	74.7	112.1	
145	27.5	22.1	33.1	




Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#90
Hexachlorobutadiene
Concen: 16.142 ug/l
RT: 13.78 min Scan# 2114
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Resp	Lower	Upper
225	100	78870		
223	63.0	0.0	123.0	
227	64.7	0.0	121.8	









#91
Naphthalene
Concen: 17.970 ug/l
RT: 13.83 min Scan# 2123
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBS01

Tot Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.2	15.2
129	11.0	8.8	13.2

Manual Integrations
APPROVED

MMDadoda
11/15/2019 9:21:03 AM



#92
1,2,3-Trichlorobenzene
Concen: 17.606 ug/l
RT: 14.01 min Scan# 2153
Delta R.T. 0.00 min
Lab File: VX013413.D
Acq: 14 Nov 2019 11:32

Tgt Ion	Ratio	Lower	Upper
180	100		
182	94.8	0.0	191.0
145	29.1	0.0	59.0

