


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX111419\
 Data File : VX013409.D
 Acq On : 14 Nov 2019 09:20
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X111119W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Tue Nov 12 01:16:13 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1674

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	30993	PASS
75	95	30	60	48.7	76933	PASS
95	95	100	100	100.0	158098	PASS
96	95	5	9	6.4	10127	PASS
173	174	0.00	2	1.2	1673	PASS
174	95	50	100	85.0	134445	PASS
175	174	5	9	7.4	9964	PASS
176	174	95	101	96.5	129775	PASS
177	176	5	9	6.2	8017	PASS