

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111422\
 Data File : VX032784.D
 Acq On : 14 Nov 2022 11:09
 Operator : JC/MD
 Sample : VX1114WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBL01

Quant Time: Nov 15 05:06:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	98455	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	160546	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	144969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68327	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67884	49.622	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.240%
35) Dibromofluoromethane	5.385	113	55496	48.460	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.920%
50) Toluene-d8	8.647	98	197354	49.560	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.120%
62) 4-Bromofluorobenzene	11.079	95	74011	47.580	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111422\
Data File : VX032784.D
Acq On : 14 Nov 2022 11:09
Operator : JC/MD
Sample : VX1114WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBL01

Quant Time: Nov 15 05:06:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

