

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111422\
 Data File : VX032801.D
 Acq On : 14 Nov 2022 17:46
 Operator : JC/MD
 Sample : N5587-12 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 INTERCEPTOR-SUMP-20221110

Quant Time: Nov 15 05:09:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	96253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	157168	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	140715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65781	49.185	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.380%		
35) Dibromofluoromethane	5.385	113	53887	48.066	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.140%		
50) Toluene-d8	8.652	98	192636	49.415	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.840%		
62) 4-Bromofluorobenzene	11.079	95	75234	49.406	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.820%		
Target Compounds						
20) Methylene Chloride	2.788	84	562	0.504	ug/l	87
44) Trichloroethene	7.122	130	523	0.421	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111422\
Data File : VX032801.D
Acq On : 14 Nov 2022 17:46
Operator : JC/MD
Sample : N5587-12 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
INTERCEPTOR-SUMP-20221110

Quant Time: Nov 15 05:09:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

