

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111422\
 Data File : VX032805.D
 Acq On : 14 Nov 2022 19:18
 Operator : JC/MD
 Sample : N5587-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-121.5N-20221110

Quant Time: Nov 15 05:10:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	96699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	159691	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66275	49.326	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.660%
35) Dibromofluoromethane	5.385	113	54321	47.688	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.380%
50) Toluene-d8	8.647	98	193573	48.871	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.740%
62) 4-Bromofluorobenzene	11.079	95	71905	46.474	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.940%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	982	2.011	ug/l #	85
20) Methylene Chloride	2.788	84	320	0.286	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111422\
Data File : VX032805.D
Acq On : 14 Nov 2022 19:18
Operator : JC/MD
Sample : N5587-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-121.5N-20221110

Quant Time: Nov 15 05:10:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

