

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111423\
 Data File : VX038797.D
 Acq On : 14 Nov 2023 11:43
 Operator : JC/MD
 Sample : VX1114WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBL01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/15/2023
 Supervised By : Mahesh Dadoda 11/15/2023

Quant Time: Nov 14 18:26:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 14 09:15:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	100575m	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	197278	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66363	46.380	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.760%
35) Dibromofluoromethane	5.385	113	60865	44.909	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.820%
50) Toluene-d8	8.647	98	236139	47.038	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.080%
62) 4-Bromofluorobenzene	11.079	95	107670	51.297	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111423\
Data File : VX038797.D
Acq On : 14 Nov 2023 11:43
Operator : JC/MD
Sample : VX1114WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBL01

Quant Time: Nov 14 18:26:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111423W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 14 09:15:55 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/15/2023
Supervised By :Mahesh Dadoda 11/15/2023

