

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111423\
 Data File : VX038807.D
 Acq On : 14 Nov 2023 15:35
 Operator : JC/MD
 Sample : 05323-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1

Quant Time: Nov 14 16:07:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 14 09:15:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	96577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	188251	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104942	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64022	46.596	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.200%
35) Dibromofluoromethane	5.385	113	58714	45.399	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.800%
50) Toluene-d8	8.647	98	228777	47.757	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.520%
62) 4-Bromofluorobenzene	11.079	95	106349	53.097	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.200%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	32130	43.641	ug/l #	85
20) Methylene Chloride	2.794	84	6350	4.611	ug/l	89
25) 2-Butanone	4.568	43	3513	3.211	ug/l	95
43) Isopropyl Acetate	6.342	43	85831	27.329	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111423\
Data File : VX038807.D
Acq On : 14 Nov 2023 15:35
Operator : JC/MD
Sample : 05323-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1

Quant Time: Nov 14 16:07:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111423W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 14 09:15:55 2023
Response via : Initial Calibration

