

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111423\
 Data File : VX038809.D
 Acq On : 14 Nov 2023 16:21
 Operator : JC/MD
 Sample : 05379-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-4

Quant Time: Nov 14 16:57:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 14 09:15:55 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	100518	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	193377	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199727	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64164	44.869	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.740%
35) Dibromofluoromethane	5.385	113	58791	44.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.500%
50) Toluene-d8	8.647	98	235250	47.807	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.620%
62) 4-Bromofluorobenzene	11.079	95	110202	53.563	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.120%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	19644	25.635	ug/l	95
18) Methyl Acetate	2.697	43	1175	0.330	ug/l #	85
20) Methylene Chloride	2.788	84	4336	3.025	ug/l #	80
43) Isopropyl Acetate	6.336	43	86609	26.846	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111423\
Data File : VX038809.D
Acq On : 14 Nov 2023 16:21
Operator : JC/MD
Sample : 05379-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-4

Quant Time: Nov 14 16:57:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111423W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 14 09:15:55 2023
Response via : Initial Calibration

