

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX111423\  
 Data File : VX038812.D  
 Acq On : 14 Nov 2023 17:30  
 Operator : JC/MD  
 Sample : 05380-08  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 WC-8

Quant Time: Nov 15 14:39:44 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X111423W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Nov 14 09:15:55 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 96575    | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 188071   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 193232   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 102045   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 63926    | 46.527   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =      | 93.060%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 57387    | 44.416   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 88.840%  |
| 50) Toluene-d8              | 8.647  | 98    | 232217   | 48.522   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =      | 97.040%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 106312   | 53.130   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 106.260% |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 2.380  | 43    | 19575    | 26.588   | ug/l # | 89       |
| 20) Methylene Chloride      | 2.788  | 84    | 4477     | 3.251    | ug/l   | 92       |
| 43) Isopropyl Acetate       | 6.336  | 43    | 83935    | 26.751   | ug/l   | 94       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX111423\  
Data File : VX038812.D  
Acq On : 14 Nov 2023 17:30  
Operator : JC/MD  
Sample : 05380-08  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
WC-8

Quant Time: Nov 15 14:39:44 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X111423W.M  
Quant Title : SW846 8260  
QLast Update : Tue Nov 14 09:15:55 2023  
Response via : Initial Calibration

