

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111425\
 Data File : VX048604.D
 Acq On : 14 Nov 2025 11:39
 Operator : JC/MD
 Sample : VX1114WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1114WBL01

Quant Time: Nov 14 22:18:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 08 05:08:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	149161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	302183	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	367753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	166817	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	121003	58.218	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.440%
35) Dibromofluoromethane	5.379	113	100133	43.781	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.560%
50) Toluene-d8	8.635	98	350615	49.153	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.061	95	148553	55.883	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111425\
Data File : VX048604.D
Acq On : 14 Nov 2025 11:39
Operator : JC/MD
Sample : VX1114WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1114WBL01

Quant Time: Nov 14 22:18:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X110725W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 08 05:08:11 2025
Response via : Initial Calibration

