

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX111518\
 Data File : VX006033.D
 Acq On : 15 Nov 2018 21:06
 Operator : JC/MD
 Sample : J5911-05 10X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FDSB-07(10-12)

Quant Time: Nov 16 09:37:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	157478	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	235142	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	190115	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	130561	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	86279	47.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.32%	
35) Dibromofluoromethane	5.50	113	71177	44.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.00%	
50) Toluene-d8	8.71	98	253495	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
62) 4-Bromofluorobenzene	11.14	95	98836	51.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.40%	
Target Compounds						
31) Cyclohexane	5.57	56	11821	4.69	ug/l	96
39) Methylcyclohexane	7.46	83	56371	24.54	ug/l #	86
73) Isopropylbenzene	11.02	105	52925	7.01	ug/l	100
78) n-propylbenzene	11.36	91	139751	15.71	ug/l	98
85) sec-Butylbenzene	11.94	105	149306	18.81	ug/l	82
89) n-Butylbenzene	12.39	91	120771	18.53	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX111518\
Data File : VX006033.D
Acq On : 15 Nov 2018 21:06
Operator : JC/MD
Sample : J5911-05 10X
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FDSB-07(10-12)

Quant Time: Nov 16 09:37:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

