

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111519\
 Data File : VX013435.D
 Acq On : 15 Nov 2019 10:41
 Operator : JC/SP
 Sample : VX1115WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1115WBL01

Quant Time: Nov 15 23:23:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	701544	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1120022	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	995495	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	441946	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	405693	48.81	ug/l	0.00
Spiked Amount						
			Recovery	=	97.62%	
35) Dibromofluoromethane	5.48	113	340885	48.53	ug/l	0.00
Spiked Amount						
			Recovery	=	97.06%	
50) Toluene-d8	8.71	98	1334645	50.40	ug/l	0.00
Spiked Amount						
			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	11.13	95	444194	46.47	ug/l	0.00
Spiked Amount						
			Recovery	=	92.94%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111519\
Data File : VX013435.D
Acq On : 15 Nov 2019 10:41
Operator : JC/SP
Sample : VX1115WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1115WBL01

Quant Time: Nov 15 23:23:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

