

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111519\
 Data File : VX013444.D
 Acq On : 15 Nov 2019 14:26
 Operator : JC/SP
 Sample : K5884-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S36-TB-2019-4

Quant Time: Nov 16 00:52:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 07:06:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	705531	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	1130043	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	1006103	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	457449	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	409199	48.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.92%	
35) Dibromofluoromethane	5.49	113	339516	47.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.80%	
50) Toluene-d8	8.71	98	1345958	50.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.74%	
62) 4-Bromofluorobenzene	11.14	95	451168	46.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111519\
Data File : VX013444.D
Acq On : 15 Nov 2019 14:26
Operator : JC/SP
Sample : K5884-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S36-TB-2019-4

Quant Time: Nov 16 00:52:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X103119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 07:06:38 2019
Response via : Initial Calibration

