

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032815.D
 Acq On : 15 Nov 2022 12:21
 Operator : JC/MD
 Sample : N5558-07DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20BR-20221109DL

Quant Time: Nov 16 03:33:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	103738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	169567	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	151731	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70291	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70148	48.666	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.340%		
35) Dibromofluoromethane	5.385	113	57570	47.597	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.200%		
50) Toluene-d8	8.647	98	209110	49.719	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.440%		
62) 4-Bromofluorobenzene	11.079	95	78795	47.961	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.920%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.167	85	490	0.410	ug/l	92
4) Vinyl Chloride	1.374	62	19652	17.647	ug/l	96
12) 1,1-Dichloroethene	2.313	96	342	0.329	ug/l #	92
20) Methylene Chloride	2.794	84	1083	0.902	ug/l	91
21) trans-1,2-Dichloroethene	3.099	96	530	0.468	ug/l #	74
27) cis-1,2-Dichloroethene	4.489	96	116533	89.027	ug/l	89
44) Trichloroethene	7.129	130	20748	15.463	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
Data File : VX032815.D
Acq On : 15 Nov 2022 12:21
Operator : JC/MD
Sample : N5558-07DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20BR-20221109DL

Quant Time: Nov 16 03:33:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

