

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032818.D
 Acq On : 15 Nov 2022 13:30
 Operator : JC/MD
 Sample : N5587-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-01-20221110

Quant Time: Nov 16 03:33:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	97825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	159739	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	142049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66734	49.096	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.200%
35) Dibromofluoromethane	5.385	113	54371	47.718	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.440%
50) Toluene-d8	8.646	98	195908	49.446	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.900%
62) 4-Bromofluorobenzene	11.079	95	74039	47.839	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.680%
Target Compounds						
					Qvalue	
16) Acetone	2.391	43	2848	5.765	ug/l	94
20) Methylene Chloride	2.788	84	2649	2.339	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
Data File : VX032818.D
Acq On : 15 Nov 2022 13:30
Operator : JC/MD
Sample : N5587-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-01-20221110

Quant Time: Nov 16 03:33:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

