

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032823.D
 Acq On : 15 Nov 2022 15:26
 Operator : JC/MD
 Sample : N5587-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 INTERCEPTOR-SUMP-20221110

Quant Time: Nov 16 03:34:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	104641	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	171827	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155222	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73074	50.259	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.520%
35) Dibromofluoromethane	5.385	113	58726	47.914	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.820%
50) Toluene-d8	8.647	98	211179	49.551	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.079	95	81955	49.228	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.460%
Target Compounds						
						Qvalue
7) Trichlorofluoromethane	1.892	101	936	0.440	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	3690	2.795	ug/l	94
44) Trichloroethene	7.129	130	9991	7.348	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
Data File : VX032823.D
Acq On : 15 Nov 2022 15:26
Operator : JC/MD
Sample : N5587-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
INTERCEPTOR-SUMP-20221110

Quant Time: Nov 16 03:34:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

