

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032841.D
 Acq On : 16 Nov 2022 03:29
 Operator : JC/MD
 Sample : N5640-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW-3

Quant Time: Nov 16 05:11:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	109791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180855	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163820	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77263	50.647	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.300%
35) Dibromofluoromethane	5.385	113	62498	48.446	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.900%
50) Toluene-d8	8.647	98	225633	50.299	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.600%
62) 4-Bromofluorobenzene	11.079	95	85389	48.731	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.460%
Target Compounds						
20) Methylene Chloride	2.788	84	1084	0.853	ug/l	Qvalue # 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
Data File : VX032841.D
Acq On : 16 Nov 2022 03:29
Operator : JC/MD
Sample : N5640-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW-3

Quant Time: Nov 16 05:11:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

