

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032843.D
 Acq On : 16 Nov 2022 04:15
 Operator : JC/MD
 Sample : N5640-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB111422

Quant Time: Nov 16 05:11:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	104669	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	174434	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	154583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70624	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74148	50.984	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.960%
35) Dibromofluoromethane	5.385	113	58035	46.642	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.280%
50) Toluene-d8	8.647	98	213926	49.445	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.880%
62) 4-Bromofluorobenzene	11.079	95	80986	47.919	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.840%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2295	4.342	ug/l	96
20) Methylene Chloride	2.794	84	1075	0.887	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
Data File : VX032843.D
Acq On : 16 Nov 2022 04:15
Operator : JC/MD
Sample : N5640-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB111422

Quant Time: Nov 16 05:11:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

