

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032846.D
 Acq On : 16 Nov 2022 05:25
 Operator : JC/MD
 Sample : N5640-13 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Nov 16 07:05:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105326	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175531	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	157840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78279	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73115	49.960	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.920%
35) Dibromofluoromethane	5.385	113	58979	47.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.200%
50) Toluene-d8	8.647	98	217562	49.971	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.940%
62) 4-Bromofluorobenzene	11.079	95	84875	49.906	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.820%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	1250	1.025	ug/l #	80
39) Methylcyclohexane	7.379	83	1757	0.852	ug/l #	87
52) Toluene	8.720	92	13139	4.082	ug/l	99
67) Ethyl Benzene	10.195	91	103765	16.412	ug/l	99
68) m/p-Xylenes	10.299	106	112331	45.754	ug/l	100
69) o-Xylene	10.640	106	20761	8.697	ug/l	99
73) Isopropylbenzene	10.963	105	16109	2.597	ug/l	99
78) n-propylbenzene	11.305	91	40249	5.632	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	67482	12.729	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	241590	45.818	ug/l	99
95) Naphthalene	13.774	128	43395	7.061	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
Data File : VX032846.D
Acq On : 16 Nov 2022 05:25
Operator : JC/MD
Sample : N5640-13 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Nov 16 07:05:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

