

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032848.D
 Acq On : 16 Nov 2022 06:11
 Operator : JC/MD
 Sample : N5640-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-119

Quant Time: Nov 16 07:05:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	99050	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	163842	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	146035	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68368	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70740	51.400	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	102.800%	
35) Dibromofluoromethane	5.385	113	55430	47.429	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.860%	
50) Toluene-d8	8.647	98	202420	49.810	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	99.620%	
62) 4-Bromofluorobenzene	11.079	95	74741	47.083	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.160%	
Target Compounds						
16) Acetone	2.380	43	1084	2.167	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
Data File : VX032848.D
Acq On : 16 Nov 2022 06:11
Operator : JC/MD
Sample : N5640-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-119

Quant Time: Nov 16 07:05:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

