

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032850.D
 Acq On : 16 Nov 2022 06:57
 Operator : JC/MD
 Sample : N5640-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-5

Quant Time: Nov 16 08:07:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	100113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	167567	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72269	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70752	50.862	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.720%
35) Dibromofluoromethane	5.385	113	56612	47.363	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.720%
50) Toluene-d8	8.647	98	207628	49.956	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.079	95	78674	48.459	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.920%
Target Compounds						
19) Methyl tert-butyl Ether	3.117	73	4581	1.298	ug/l	91
30) Chloroform	5.093	83	1210	0.554	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
Data File : VX032850.D
Acq On : 16 Nov 2022 06:57
Operator : JC/MD
Sample : N5640-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-5

Quant Time: Nov 16 08:07:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

