

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032851.D
 Acq On : 16 Nov 2022 07:20
 Operator : JC/MD
 Sample : N5640-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6

Quant Time: Nov 16 08:08:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	101197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	171121	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153985	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74693	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71821	51.078	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.160%
35) Dibromofluoromethane	5.385	113	57077	46.761	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.520%
50) Toluene-d8	8.647	98	211489	49.828	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.660%
62) 4-Bromofluorobenzene	11.079	95	81878	49.385	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
Data File : VX032851.D
Acq On : 16 Nov 2022 07:20
Operator : JC/MD
Sample : N5640-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Nov 16 08:08:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

