

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032835.D
 Acq On : 15 Nov 2022 20:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 16 03:36:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.577	0.510	11.6	66	0.00
3 P	Chloromethane	0.488	0.400	18.0	64	0.00
4 C	Vinyl Chloride	0.537	0.473	11.9#	67	0.00
5 T	Bromomethane	0.429	0.394	8.2	71	0.01
6 T	Chloroethane	0.360	0.466	-29.4#	109	0.00
7 T	Trichlorofluoromethane	1.017	0.963	5.3	75	0.00
8 T	Diethyl Ether	0.330	0.321	2.7	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.537	0.542	-0.9	82	0.00
10 T	Methyl Iodide	0.558	0.578	-3.6	77	0.00
11 T	Tert butyl alcohol	0.101	0.096	5.0	76	0.04
12 CM	1,1-Dichloroethene	0.501	0.481	4.0#	77	0.00
13 T	Acrolein	0.101	0.094	6.9	78	0.00
14 T	Allyl chloride	0.799	0.790	1.1	79	0.00
15 T	Acrylonitrile	0.274	0.257	6.2	78	0.00
16 T	Acetone	0.253	0.221	12.6	76	0.01
17 T	Carbon Disulfide	1.194	0.952	20.3	62	0.00
18 T	Methyl Acetate	0.821	0.855	-4.1	84	0.00
19 T	Methyl tert-butyl Ether	1.762	1.848	-4.9	87	0.00
20 T	Methylene Chloride	0.579	0.563	2.8	83	0.00
21 T	trans-1,2-Dichloroethene	0.546	0.527	3.5	79	0.00
22 T	Diisopropyl ether	1.729	1.888	-9.2	90	0.00
23 T	Vinyl Acetate	1.397	1.394	0.2	80	0.00
24 P	1,1-Dichloroethane	0.994	1.037	-4.3	85	0.00
25 T	2-Butanone	0.391	0.353	9.7	75	0.00
26 T	2,2-Dichloropropane	0.791	0.808	-2.1	82	0.00
27 T	cis-1,2-Dichloroethene	0.631	0.672	-6.5	85	0.00
28 T	Bromochloromethane	0.354	0.323	8.8	78	0.00
29 T	Tetrahydrofuran	0.249	0.222	10.8	73	0.01
30 C	Chloroform	1.091	1.195	-9.5#	90	0.00
31 T	Cyclohexane	0.842	0.792	5.9	74	0.00
32 T	1,1,1-Trichloroethane	0.979	1.052	-7.5	86	0.00
33 S	1,2-Dichloroethane-d4	0.695	0.691	0.6	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.357	0.364	-2.0	86	0.00
36 T	1,1-Dichloropropene	0.497	0.494	0.6	81	0.00
37 T	Ethyl Acetate	0.517	0.459	11.2	77	0.00
38 T	Carbon Tetrachloride	0.551	0.571	-3.6	83	0.00
39 T	Methylcyclohexane	0.588	0.545	7.3	73	0.00
40 TM	Benzene	1.405	1.443	-2.7	84	0.00
41 T	Methacrylonitrile	0.263	0.264	-0.4	82	0.00
42 TM	1,2-Dichloroethane	0.575	0.609	-5.9	87	0.00
43 T	Isopropyl Acetate	0.794	0.772	2.8	78	0.00
44 TM	Trichloroethene	0.396	0.398	-0.5	83	0.00
45 C	1,2-Dichloropropane	0.361	0.377	-4.4#	86	0.00
46 T	Dibromomethane	0.277	0.285	-2.9	85	0.00
47 T	Bromodichloromethane	0.524	0.583	-11.3	90	0.00
48 T	Methyl methacrylate	0.410	0.401	2.2	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032835.D
 Acq On : 15 Nov 2022 20:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 16 03:36:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	71	0.05
50 S	Toluene-d8	1.240	1.218	1.8	81	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.488	2.4	79	0.00
52 CM	Toluene	0.917	0.956	-4.3#	84	0.00
53 T	t-1,3-Dichloropropene	0.535	0.570	-6.5	83	0.00
54 T	cis-1,3-Dichloropropene	0.572	0.623	-8.9	86	0.00
55 T	1,1,2-Trichloroethane	0.386	0.415	-7.5	90	0.00
56 T	Ethyl methacrylate	0.556	0.586	-5.4	82	0.00
57 T	1,3-Dichloropropane	0.624	0.663	-6.3	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.268	0.266	0.7	78	0.00
59 T	2-Hexanone	0.371	0.348	6.2	75	0.00
60 T	Dibromochloromethane	0.422	0.474	-12.3	91	0.00
61 T	1,2-Dibromoethane	0.409	0.429	-4.9	85	0.00
62 S	4-Bromofluorobenzene	0.484	0.514	-6.2	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.366	0.376	-2.7	84	0.00
65 PM	Chlorobenzene	1.120	1.180	-5.4	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.422	0.464	-10.0	90	0.00
67 C	Ethyl Benzene	2.003	2.145	-7.1#	86	0.00
68 T	m/p-Xylenes	0.778	0.819	-5.3	84	0.00
69 T	o-Xylene	0.756	0.820	-8.5	87	0.00
70 T	Styrene	1.257	1.385	-10.2	87	0.00
71 P	Bromoform	0.333	0.365	-9.6	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.962	4.080	-3.0	87	0.00
74 T	N-amyl acetate	1.497	1.420	5.1	77	0.00
75 P	1,1,2,2-Tetrachloroethane	1.272	1.249	1.8	84	0.00
76 T	1,2,3-Trichloropropane	1.112	1.103	0.8	82	0.00
77 T	Bromobenzene	0.971	1.004	-3.4	89	0.00
78 T	n-propylbenzene	4.565	4.681	-2.5	86	0.00
79 T	2-Chlorotoluene	2.734	2.822	-3.2	87	0.00
80 T	1,3,5-Trimethylbenzene	3.386	3.503	-3.5	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.350	0.326	6.9	75	0.00
82 T	4-Chlorotoluene	3.199	3.324	-3.9	87	0.00
83 T	tert-Butylbenzene	3.381	3.561	-5.3	88	0.00
84 T	1,2,4-Trimethylbenzene	3.368	3.500	-3.9	85	0.00
85 T	sec-Butylbenzene	4.068	4.248	-4.4	85	0.00
86 T	p-Isopropyltoluene	3.452	3.599	-4.3	84	0.00
87 T	1,3-Dichlorobenzene	1.832	1.904	-3.9	88	0.00
88 T	1,4-Dichlorobenzene	1.901	1.934	-1.7	87	0.00
89 T	n-Butylbenzene	3.003	3.103	-3.3	83	0.00
90 T	Hexachloroethane	0.564	0.587	-4.1	85	0.00
91 T	1,2-Dichlorobenzene	1.807	1.900	-5.1	88	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.249	10.8	73	0.00
93 T	1,2,4-Trichlorobenzene	1.178	1.175	0.3	86	0.00
94 T	Hexachlorobutadiene	0.492	0.494	-0.4	87	0.00
95 T	Naphthalene	3.925	3.782	3.6	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
Data File : VX032835.D
Acq On : 15 Nov 2022 20:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 16 03:36:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.197	1.169	2.3	84	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6