



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111522\
 Data File : VX032809.D
 Acq On : 15 Nov 2022 08:29
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Title : SW846 8260
 Last Update : Thu Oct 27 08:28:06 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	3550	PASS
75	95	30	60	55.7	10270	PASS
95	95	100	100	100.0	18449	PASS
96	95	5	9	7.1	1305	PASS
173	174	0.00	2	1.1	175	PASS
174	95	50	100	82.5	15224	PASS
175	174	5	9	7.8	1194	PASS
176	174	95	101	95.7	14572	PASS
177	176	5	9	6.6	962	PASS