

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111524\
 Data File : VX043869.D
 Acq On : 15 Nov 2024 12:36
 Operator : JC/MD
 Sample : P4785-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E27P0

Quant Time: Nov 16 00:11:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML111424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 15 03:21:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	210395	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	184255	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	80887	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	64576	49.430	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.860%
7) Chloroethane-d5	1.642	69	48397	54.978	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.960%
11) 1,1-Dichloroethene-d2	2.294	65	32363	54.905	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	109.820%
21) 2-Butanone-d5	4.459	46	77963	112.223	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.220%
24) Chloroform-d	5.050	84	148069	53.271	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.540%
26) 1,2-Dichloroethane-d4	5.952	65	100072	56.378	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.760%
32) Benzene-d6	5.964	84	284921	55.580	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.160%
36) 1,2-Dichloropropane-d6	7.306	67	81993	56.160	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	112.320%
41) Toluene-d8	8.647	98	265086	54.688	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.380%
43) trans-1,3-Dichloroprop...	8.946	79	38765	52.685	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	105.360%
47) 2-Hexanone-d5	9.384	63	61102	113.613	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.610%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	108304	53.911	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.820%
66) 1,2-Dichlorobenzene-d4	12.317	152	95233	56.437	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111524\
Data File : VX043869.D
Acq On : 15 Nov 2024 12:36
Operator : JC/MD
Sample : P4785-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E27P0

Quant Time: Nov 16 00:11:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM111424WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 15 03:21:47 2024
Response via : Initial Calibration

