

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111524\
 Data File : VX043880.D
 Acq On : 15 Nov 2024 16:50
 Operator : JC/MD
 Sample : P4785-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E27P3

Quant Time: Nov 16 00:12:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML111424WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 15 03:21:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	210206	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	188804	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	83759	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	58453	44.784	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	89.560%	
7) Chloroethane-d5	1.648	69	43101	49.006	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	98.020%	
11) 1,1-Dichloroethene-d2	2.294	65	29427	49.969	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	99.940%	
21) 2-Butanone-d5	4.465	46	75016	108.078	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	108.080%	
24) Chloroform-d	5.056	84	137502	49.514	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.020%	
26) 1,2-Dichloroethane-d4	5.952	65	93799	52.891	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	105.780%	
32) Benzene-d6	5.964	84	266166	50.670	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.340%	
36) 1,2-Dichloropropane-d6	7.306	67	77976	52.122	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	104.240%	
41) Toluene-d8	8.647	98	247003	49.730	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	99.460%	
43) trans-1,3-Dichloroprop...	8.952	79	35779	47.455	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	94.920%	
47) 2-Hexanone-d5	9.384	63	60766	110.266	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	110.270%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	107897	52.414	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	104.820%	
66) 1,2-Dichlorobenzene-d4	12.317	152	92897	53.165	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	106.320%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111524\
Data File : VX043880.D
Acq On : 15 Nov 2024 16:50
Operator : JC/MD
Sample : P4785-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E27P3

Quant Time: Nov 16 00:12:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM111424WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 15 03:21:47 2024
Response via : Initial Calibration

