



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111524\
Data File : VX043859.D
Acq On : 15 Nov 2024 08:03
Operator : JC/MD
Sample : BFB700
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB700

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM111424WMA.M
Title : VOC Analysis
Last Update : Fri Nov 15 03:21:47 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	4410	PASS
75	95	30	80	47.5	13018	PASS
95	95	100	100	100.0	27384	PASS
96	95	5	9	5.4	1470	PASS
173	174	0.00	2	1.0	212	PASS
174	95	50	120	79.8	21861	PASS
175	174	5	9	7.7	1675	PASS
176	174	95	101	97.3	21267	PASS
177	176	5	9	6.7	1416	PASS

Scan 1633 (11.043 min): VX043859.D\data.ms
BFB700

m/z	Abundance
40.10	336.0
44.10	99.0

Instrument :
MSVOA_X
ClientSampleId :
BFB700

m/z	Abundance
36.10	192.0
37.10	1179.0
38.00	1011.0
39.10	459.0
40.20	320.0
44.00	183.0
45.10	158.0
47.20	281.0
47.80	86.0
49.00	905.0
50.10	4635.0
51.00	1316.0
55.10	86.0
56.10	388.0
57.10	627.0
60.20	274.0
61.00	1225.0
62.00	1414.0
63.00	953.0
64.00	113.0
67.00	51.0
68.00	2262.0
69.00	2294.0
69.90	250.0
72.10	109.0
73.00	1131.0
74.10	4561.0
75.00	13089.0
76.10	1004.0
76.80	66.0
78.10	117.0
79.00	1147.0
79.90	220.0
80.90	1126.0
82.00	242.0
86.90	1226.0
87.80	975.0
90.80	72.0
92.10	911.0
93.00	965.0
94.00	3636.0
95.10	26128.0
96.00	1394.0
104.00	91.0
106.00	143.0
115.90	213.0
117.10	271.0
118.10	95.0
118.90	155.0
128.00	142.0
129.90	69.0
135.00	79.0
141.00	303.0
143.10	344.0
148.00	54.0
174.00	17328.0
175.00	1176.0
176.00	16552.0
177.00	963.0
178.30	64.0

Instrument :
MSVOA_X
ClientSampleId :
BFB700

m/z	Abundance
36.10	203.0
37.00	1273.0
38.20	1003.0
39.20	444.0
39.90	331.0
43.90	341.0
44.90	235.0
46.90	258.0
47.80	75.0
49.00	1199.0
50.10	5104.0
51.10	1337.0
55.10	129.0
56.10	481.0
57.10	982.0
60.10	549.0
60.90	1184.0
62.10	1791.0
63.00	1078.0
63.90	64.0
68.10	2671.0
69.10	3226.0
69.90	246.0
71.90	161.0
73.10	1271.0
74.10	4766.0
75.10	15717.0
76.00	1322.0
76.70	76.0
78.10	153.0
78.90	1217.0
80.00	312.0
81.00	1378.0
82.10	313.0
87.00	1030.0
88.00	1204.0
90.80	180.0
91.90	885.0
93.00	1439.0
94.00	3665.0
95.10	32584.0
96.10	1768.0
103.90	153.0
105.90	266.0
109.90	66.0
111.20	65.0
115.90	127.0
116.90	379.0
117.90	149.0
118.90	240.0
127.90	191.0
130.00	101.0
135.00	69.0
140.90	327.0
143.00	613.0
145.00	71.0
146.10	65.0
157.20	80.0
172.10	172.0
173.00	348.0
174.00	26320.0
175.10	1976.0
176.10	24600.0
177.00	1702.0

Instrument :
MSVOA_X
ClientSampleId :
BFB700

m/z	Abundance
36.10	154.0
37.10	727.0
38.00	987.0
39.10	442.0
40.00	291.0
44.20	190.0
44.90	99.0
46.70	113.0
47.40	83.0
48.20	97.0
49.00	865.0
50.00	3490.0
51.00	991.0
56.00	375.0
57.20	572.0
60.00	238.0
61.00	1046.0
62.10	1397.0
63.10	593.0
68.00	1854.0
69.10	1937.0
69.90	195.0
70.20	188.0
71.80	160.0
73.00	901.0
74.10	3564.0
75.00	10248.0
76.10	641.0
79.00	1251.0
80.00	261.0
80.90	834.0
82.00	232.0
86.90	844.0
88.10	883.0
90.90	134.0
92.00	770.0
93.10	895.0
94.00	2966.0
95.10	23440.0
96.00	1249.0
104.10	178.0
106.00	97.0
110.00	58.0
110.70	74.0
115.90	142.0
116.80	193.0
117.80	147.0
128.00	180.0
130.00	256.0
131.10	72.0
141.00	410.0
142.80	400.0
145.10	67.0
145.70	51.0
146.10	76.0
157.10	58.0
172.10	90.0
172.80	289.0
174.00	21936.0
175.10	1872.0
176.10	22648.0
177.10	1584.0

Instrument :
MSVOA_X
ClientSampleId :
BFB700