

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111618\
 Data File : VX006039.D
 Acq On : 16 Nov 2018 10:27
 Operator : JC/MD
 Sample : VX1116WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1116WBL01

Quant Time: Nov 17 01:14:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	146283	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	213997	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	202647	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	111673	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	84314	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
35) Dibromofluoromethane	5.51	113	66451	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.32%	
50) Toluene-d8	8.71	98	224926	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
62) 4-Bromofluorobenzene	11.14	95	94624	54.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.78%	
Target Compounds						
16) Acetone	2.45	43	1753	1.936	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111618\
Data File : VX006039.D
Acq On : 16 Nov 2018 10:27
Operator : JC/MD
Sample : VX1116WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX1116WBL01

Quant Time: Nov 17 01:14:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

