

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111618\
 Data File : VX006046.D
 Acq On : 16 Nov 2018 13:35
 Operator : JC/MD
 Sample : J5934-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S36-0149

Manual Integrations
 APPROVED

MMDadoda
 11/19/2018 12:22:29 PM

Quant Time: Nov 17 01:27:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 07 14:40:53 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	139048	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.86	114	196634	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	183287	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	99549	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	80042	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
35) Dibromofluoromethane	5.51	113	65884	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	8.71	98	215259	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
62) 4-Bromofluorobenzene	11.14	95	77648	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
Target Compounds						
4) Vinyl Chloride	1.41	62	47165	28.555	ug/l	100
16) Acetone	2.45	43	1354	1.573	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	5482	4.229	ug/l	93
27) cis-1,2-Dichloroethene	4.59	96	14496	10.127	ug/l	89
40) Benzene	6.14	78	13339	2.414	ug/l	93
65) Chlorobenzene	10.14	112	115260	31.390	ug/l	100
88) 1,4-Dichlorobenzene	12.10	146	3457m	1.084	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111618\
Data File : VX006046.D
Acq On : 16 Nov 2018 13:35
Operator : JC/MD
Sample : J5934-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S36-0149

Manual Integrations
APPROVED

MMDadoda
11/19/2018 12:22:29 PM

Quant Time: Nov 17 01:27:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X110718W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 07 14:40:53 2018
Response via : Initial Calibration

