

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111620\
 Data File : VX019430.D
 Acq On : 16 Nov 2020 13:02
 Operator : JC/MD
 Sample : VSTDCCC0.5
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTD0.5209

Quant Time: Nov 17 09:10:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM111320.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 17 04:26:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.82	114	19556	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.09	117	16586	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.06	152	7786	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.39	65	10216	0.50	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	100.00%
4) 1,2-Dichloroethane-d4	6.01	65	8373	0.46	ug/L	-0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.00%
7) 1,2-Dichloropropane-d6	7.35	67	8533	0.56	ug/L	-0.02
Spiked Amount	0.500	Range	60 - 140	Recovery	=	112.00%
8) Toluene-d8	8.68	98	22612	0.53	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.00%
10) 1,1,2,2-Tetrachloroethane-	11.22	84	5983	0.56	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	112.00%

Target Compounds

					Qvalue
3) Vinyl chloride	1.39	62	8554	0.450	ug/L 98
6) Trichloroethene	7.18	95	6856	0.500	ug/L 95
9) 1,2-Dibromoethane	9.64	107	4000	0.495	ug/L 94
12) 1,2,3-Trichloropropane	11.27	75	3810	0.495	ug/L 99
13) 1,2-Dibromo-3-chloropropan	12.97	75	656	0.496	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111620\
Data File : VX019430.D
Acq On : 16 Nov 2020 13:02
Operator : JC/MD
Sample : VSTDCCC0.5
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTD0.5209

Quant Time: Nov 17 09:10:04 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM111320.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 17 04:26:19 2020
Response via : Initial Calibration

