

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111620\
 Data File : VX019433.D
 Acq On : 16 Nov 2020 14:28
 Operator : JC/MD
 Sample : VX1116WBL05
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK205

Quant Time: Nov 17 09:12:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM111320.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 17 04:26:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.82	114	15419	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.09	117	13099	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	12.07	152	5713	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.39	65	7901	0.49	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	98.00%
4) 1,2-Dichloroethane-d4	6.03	65	6542	0.45	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	90.00%
7) 1,2-Dichloropropane-d6	7.37	67	6673	0.56	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	112.00%
8) Toluene-d8	8.69	98	17247	0.52	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.00%
10) 1,1,2,2-Tetrachloroethane-	11.23	84	4390	0.52	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	104.00%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111620\
Data File : VX019433.D
Acq On : 16 Nov 2020 14:28
Operator : JC/MD
Sample : VX1116WBL05
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK205

Quant Time: Nov 17 09:12:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM111320.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 17 04:26:19 2020
Response via : Initial Calibration

