

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX111620\
 Data File : VX019437.D
 Acq On : 16 Nov 2020 16:31
 Operator : JC/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTD0.5210

Quant Time: Nov 17 09:10:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM111320.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 17 04:26:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.82	114	16843	0.50	ug/L	0.00
5) Chlorobenzene-d5	10.10	117	16386	0.50	ug/L	0.02
11) 1,4-Dichlorobenzene-d4	12.06	152	10679	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.39	65	7719	0.44	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	88.00%
4) 1,2-Dichloroethane-d4	6.03	65	6836	0.43	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	86.00%
7) 1,2-Dichloropropane-d6	7.37	67	6777	0.45	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	90.00%
8) Toluene-d8	8.68	98	19963	0.48	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.00%
10) 1,1,2,2-Tetrachloroethane-	11.22	84	4120	0.39	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	78.00%

Target Compounds

					Qvalue
3) Vinyl chloride	1.39	62	6780	0.414	ug/L 96
6) Trichloroethene	7.18	95	5678	0.419	ug/L 96
9) 1,2-Dibromoethane	9.66	107	3081	0.386	ug/L 98
12) 1,2,3-Trichloropropane	11.27	75	2923	0.277	ug/L 96
13) 1,2-Dibromo-3-chloropropan	12.97	75	458	0.253	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX111620\
Data File : VX019437.D
Acq On : 16 Nov 2020 16:31
Operator : JC/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTD0.5210

Quant Time: Nov 17 09:10:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXSIM111320.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 17 04:26:19 2020
Response via : Initial Calibration

