

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
 Data File : VX032870.D
 Acq On : 16 Nov 2022 16:01
 Operator : JC/MD
 Sample : N5614-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-1

Quant Time: Nov 17 05:43:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	101549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	166873	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	149549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	68029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71575	50.726	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.460%
35) Dibromofluoromethane	5.385	113	55819	46.894	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.780%
50) Toluene-d8	8.647	98	207358	50.098	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.200%
62) 4-Bromofluorobenzene	11.079	95	78191	48.362	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.720%
Target Compounds						
19) Methyl tert-butyl Ether	3.111	73	1757	0.491	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
Data File : VX032870.D
Acq On : 16 Nov 2022 16:01
Operator : JC/MD
Sample : N5614-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-1

Quant Time: Nov 17 05:43:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

