

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
 Data File : VX032871.D
 Acq On : 16 Nov 2022 16:24
 Operator : JC/MD
 Sample : N5614-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-2

Quant Time: Nov 17 05:43:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	99137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	165874	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	148585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69676	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69834	50.697	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.400%	
35) Dibromofluoromethane	5.391	113	55437	46.854	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.700%	
50) Toluene-d8	8.647	98	203593	49.485	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.980%	
62) 4-Bromofluorobenzene	11.079	95	77737	48.371	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.740%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
Data File : VX032871.D
Acq On : 16 Nov 2022 16:24
Operator : JC/MD
Sample : N5614-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-2

Quant Time: Nov 17 05:43:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

