

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
 Data File : VX032873.D
 Acq On : 16 Nov 2022 17:10
 Operator : JC/MD
 Sample : N5640-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW-3

Quant Time: Nov 17 05:43:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	97244	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	163029	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	146687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66292	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70365	52.077	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.160%	
35) Dibromofluoromethane	5.385	113	55461	47.692	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.380%	
50) Toluene-d8	8.647	98	203364	50.292	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.580%	
62) 4-Bromofluorobenzene	11.079	95	76928	48.702	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.400%	
Target Compounds						Qvalue
20) Methylene Chloride	2.794	84	931	0.827	ug/l	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
Data File : VX032873.D
Acq On : 16 Nov 2022 17:10
Operator : JC/MD
Sample : N5640-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW-3

Quant Time: Nov 17 05:43:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

