

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
 Data File : VX032876.D
 Acq On : 16 Nov 2022 18:20
 Operator : JC/MD
 Sample : PB149082TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149082TB

Quant Time: Nov 17 05:44:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	98658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	167285	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153689	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71460	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72429	52.836	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.680%	
35) Dibromofluoromethane	5.391	113	56445	47.303	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.600%	
50) Toluene-d8	8.647	98	210290	50.682	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.360%	
62) 4-Bromofluorobenzene	11.079	95	81585	50.337	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.680%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	10535	21.145	ug/l	98
20) Methylene Chloride	2.794	84	6986	6.115	ug/l	92
37) Ethyl Acetate	4.715	43	15394	8.900	ug/l	98
43) Isopropyl Acetate	6.336	43	69911	26.325	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
Data File : VX032876.D
Acq On : 16 Nov 2022 18:20
Operator : JC/MD
Sample : PB149082TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149082TB

Quant Time: Nov 17 05:44:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

