

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
 Data File : VX032880.D
 Acq On : 16 Nov 2022 19:52
 Operator : JC/MD
 Sample : N5627-02 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WAMR-SILICA-1115-B

Quant Time: Nov 17 05:44:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	130098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	219499	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	203615	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87560	48.438	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.880%		
35) Dibromofluoromethane	5.391	113	74528	47.600	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.200%		
50) Toluene-d8	8.647	98	271744	49.913	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.820%		
62) 4-Bromofluorobenzene	11.079	95	116791	54.917	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	109.840%		
Target Compounds					Qvalue	
11) Tert butyl alcohol	2.971	59	87519	332.638	ug/l #	89
16) Acetone	2.386	43	300466	457.325	ug/l	99
18) Methyl Acetate	2.703	43	36701	17.178	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	80242	17.498	ug/l	96
20) Methylene Chloride	2.788	84	2715	1.802	ug/l	96
25) 2-Butanone	4.562	43	115006	112.985	ug/l	97
40) Benzene	6.043	78	21069	3.415	ug/l	100
43) Isopropyl Acetate	6.342	43	15228	4.370	ug/l	97
51) 4-Methyl-2-Pentanone	8.573	43	17081	7.782	ug/l #	84
52) Toluene	8.720	92	344743	85.641	ug/l	99
67) Ethyl Benzene	10.195	91	464633	56.966	ug/l	99
68) m/p-Xylenes	10.299	106	890431	281.147	ug/l	97
69) o-Xylene	10.640	106	538875	174.998	ug/l	97
73) Isopropylbenzene	10.963	105	53805	6.207	ug/l	100
78) n-propylbenzene	11.305	91	137602	13.778	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	246523	33.277	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	1042017	141.420	ug/l	98
85) sec-Butylbenzene	11.890	105	13182m	1.481	ug/l	
86) p-Isopropyltoluene	12.012	119	9767	1.293	ug/l #	73
89) n-Butylbenzene	12.323	91	22422	3.413	ug/l #	19
95) Naphthalene	13.774	128	133217	15.513	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
Data File : VX032880.D
Acq On : 16 Nov 2022 19:52
Operator : JC/MD
Sample : N5627-02 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WAMR-SILICA-1115-B

Quant Time: Nov 17 05:44:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

