

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
 Data File : VX032882.D
 Acq On : 16 Nov 2022 20:38
 Operator : JC/MD
 Sample : N5650-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 34270

Quant Time: Nov 17 05:45:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	139075	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226571	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201300	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98384	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92045	47.632	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.260%
35) Dibromofluoromethane	5.391	113	77920	48.213	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.420%
50) Toluene-d8	8.647	98	279248	49.691	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.380%
62) 4-Bromofluorobenzene	11.079	95	110752	50.452	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.900%
Target Compounds						
					Qvalue	
69) o-Xylene	10.640	106	4494	1.476	ug/l	94
84) 1,2,4-Trimethylbenzene	11.756	105	10734	1.620	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
Data File : VX032882.D
Acq On : 16 Nov 2022 20:38
Operator : JC/MD
Sample : N5650-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
34270

Quant Time: Nov 17 05:45:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

