

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
 Data File : VX032884.D
 Acq On : 16 Nov 2022 21:25
 Operator : JC/MD
 Sample : N5650-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 34267

Quant Time: Nov 17 08:33:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205950	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182586	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86855	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87129	50.245	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.500%
35) Dibromofluoromethane	5.385	113	69636	47.402	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.800%
50) Toluene-d8	8.647	98	252761	49.481	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.960%
62) 4-Bromofluorobenzene	11.079	95	98583	49.405	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.820%
Target Compounds						Qvalue
20) Methylene Chloride	2.794	84	1070	0.740	ug/l #	83
52) Toluene	8.720	92	2198	0.582	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
Data File : VX032884.D
Acq On : 16 Nov 2022 21:25
Operator : JC/MD
Sample : N5650-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
34267

Quant Time: Nov 17 08:33:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

