

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
 Data File : VX032859.D
 Acq On : 16 Nov 2022 11:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 17 05:41:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.577	0.540	6.4	70	0.00
3 P	Chloromethane	0.488	0.429	12.1	68	0.00
4 C	Vinyl Chloride	0.537	0.502	6.5#	71	0.00
5 T	Bromomethane	0.429	0.463	-7.9	84	0.01
6 T	Chloroethane	0.360	0.475	-31.9#	111	0.00
7 T	Trichlorofluoromethane	1.017	1.038	-2.1	81	0.00
8 T	Diethyl Ether	0.330	0.335	-1.5	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.537	0.588	-9.5	89	0.00
10 T	Methyl Iodide	0.558	0.593	-6.3	79	0.00
11 T	Tert butyl alcohol	0.101	0.092	8.9	73	0.02
12 CM	1,1-Dichloroethene	0.501	0.510	-1.8#	82	0.00
13 T	Acrolein	0.101	0.116	-14.9	96	0.00
14 T	Allyl chloride	0.799	0.846	-5.9	84	0.00
15 T	Acrylonitrile	0.274	0.266	2.9	81	0.00
16 T	Acetone	0.253	0.303	-19.8	104	0.00
17 T	Carbon Disulfide	1.194	0.995	16.7	65	0.00
18 T	Methyl Acetate	0.821	0.888	-8.2	87	0.00
19 T	Methyl tert-butyl Ether	1.762	1.970	-11.8	92	0.00
20 T	Methylene Chloride	0.579	0.598	-3.3	89	0.00
21 T	trans-1,2-Dichloroethene	0.546	0.556	-1.8	84	0.00
22 T	Diisopropyl ether	1.729	2.019	-16.8	96	0.00
23 T	Vinyl Acetate	1.397	1.481	-6.0	85	0.00
24 P	1,1-Dichloroethane	0.994	1.102	-10.9	90	0.00
25 T	2-Butanone	0.391	0.397	-1.5	84	0.00
26 T	2,2-Dichloropropane	0.791	0.917	-15.9	93	0.00
27 T	cis-1,2-Dichloroethene	0.631	0.719	-13.9	91	0.00
28 T	Bromochloromethane	0.354	0.345	2.5	83	-0.01
29 T	Tetrahydrofuran	0.249	0.233	6.4	76	0.00
30 C	Chloroform	1.091	1.281	-17.4#	96	0.00
31 T	Cyclohexane	0.842	0.858	-1.9	80	0.00
32 T	1,1,1-Trichloroethane	0.979	1.146	-17.1	93	0.00
33 S	1,2-Dichloroethane-d4	0.695	0.729	-4.9	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.357	0.395	-10.6	91	0.00
36 T	1,1-Dichloropropene	0.497	0.535	-7.6	86	0.00
37 T	Ethyl Acetate	0.517	0.492	4.8	81	0.00
38 T	Carbon Tetrachloride	0.551	0.623	-13.1	89	-0.01
39 T	Methylcyclohexane	0.588	0.607	-3.2	80	0.00
40 TM	Benzene	1.405	1.553	-10.5	88	0.00
41 T	Methacrylonitrile	0.263	0.286	-8.7	87	0.00
42 TM	1,2-Dichloroethane	0.575	0.658	-14.4	92	0.00
43 T	Isopropyl Acetate	0.794	0.835	-5.2	83	0.00
44 TM	Trichloroethene	0.396	0.441	-11.4	90	0.00
45 C	1,2-Dichloropropane	0.361	0.416	-15.2#	93	0.00
46 T	Dibromomethane	0.277	0.303	-9.4	88	0.00
47 T	Bromodichloromethane	0.524	0.627	-19.7	95	0.00
48 T	Methyl methacrylate	0.410	0.432	-5.4	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
 Data File : VX032859.D
 Acq On : 16 Nov 2022 11:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 17 05:41:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	69	0.02
50 S	Toluene-d8	1.240	1.301	-4.9	85	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.519	-3.8	82	0.00
52 CM	Toluene	0.917	1.043	-13.7#	90	0.00
53 T	t-1,3-Dichloropropene	0.535	0.642	-20.0	92	0.00
54 T	cis-1,3-Dichloropropene	0.572	0.685	-19.8	93	0.00
55 T	1,1,2-Trichloroethane	0.386	0.449	-16.3	96	0.00
56 T	Ethyl methacrylate	0.556	0.632	-13.7	87	0.00
57 T	1,3-Dichloropropane	0.624	0.722	-15.7	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.268	0.318	-18.7	91	0.00
59 T	2-Hexanone	0.371	0.394	-6.2	83	0.00
60 T	Dibromochloromethane	0.422	0.508	-20.4	95	0.00
61 T	1,2-Dibromoethane	0.409	0.469	-14.7	91	0.00
62 S	4-Bromofluorobenzene	0.484	0.546	-12.8	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.366	0.412	-12.6	90	0.00
65 PM	Chlorobenzene	1.120	1.294	-15.5	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.422	0.509	-20.6	96	0.00
67 C	Ethyl Benzene	2.003	2.333	-16.5#	92	0.00
68 T	m/p-Xylenes	0.778	0.899	-15.6	91	0.00
69 T	o-Xylene	0.756	0.890	-17.7	92	0.00
70 T	Styrene	1.257	1.514	-20.4	94	0.00
71 P	Bromoform	0.333	0.390	-17.1	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.962	4.415	-11.4	94	0.00
74 T	N-amyl acetate	1.497	1.562	-4.3	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.272	1.314	-3.3	88	0.00
76 T	1,2,3-Trichloropropane	1.112	1.044	6.1	77	0.00
77 T	Bromobenzene	0.971	1.082	-11.4	95	0.00
78 T	n-propylbenzene	4.565	5.114	-12.0	93	0.00
79 T	2-Chlorotoluene	2.734	3.059	-11.9	94	0.00
80 T	1,3,5-Trimethylbenzene	3.386	3.822	-12.9	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.350	0.356	-1.7	81	0.00
82 T	4-Chlorotoluene	3.199	3.590	-12.2	93	0.00
83 T	tert-Butylbenzene	3.381	3.799	-12.4	93	0.00
84 T	1,2,4-Trimethylbenzene	3.368	3.729	-10.7	90	0.00
85 T	sec-Butylbenzene	4.068	4.633	-13.9	92	0.00
86 T	p-Isopropyltoluene	3.452	3.984	-15.4	92	0.00
87 T	1,3-Dichlorobenzene	1.832	2.059	-12.4	94	0.00
88 T	1,4-Dichlorobenzene	1.901	2.109	-10.9	95	0.00
89 T	n-Butylbenzene	3.003	3.452	-15.0	91	0.00
90 T	Hexachloroethane	0.564	0.634	-12.4	91	0.00
91 T	1,2-Dichlorobenzene	1.807	2.047	-13.3	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.264	5.4	76	0.00
93 T	1,2,4-Trichlorobenzene	1.178	1.278	-8.5	93	0.00
94 T	Hexachlorobutadiene	0.492	0.565	-14.8	99	0.00
95 T	Naphthalene	3.925	4.017	-2.3	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111622\
Data File : VX032859.D
Acq On : 16 Nov 2022 11:39
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 17 05:41:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.197	1.283	-7.2	91	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6