

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
 Data File : VX038843.D
 Acq On : 16 Nov 2023 17:26
 Operator : JC/MD
 Sample : VX1116WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1116WBL02

Quant Time: Nov 16 23:04:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	133279	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	270646	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	275988	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	97659	45.474	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.940%
35) Dibromofluoromethane	5.385	113	86091	44.044	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.080%
50) Toluene-d8	8.647	98	348281	46.570	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.140%
62) 4-Bromofluorobenzene	11.079	95	146546	48.950	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
Data File : VX038843.D
Acq On : 16 Nov 2023 17:26
Operator : JC/MD
Sample : VX1116WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1116WBL02

Quant Time: Nov 16 23:04:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

