

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
 Data File : VX038847.D
 Acq On : 16 Nov 2023 18:58
 Operator : JC/MD
 Sample : 05241-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-4-20231102

Quant Time: Nov 16 23:06:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135250	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	278035	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	284246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142411	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	100264	46.006	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.020%
35) Dibromofluoromethane	5.385	113	88403	44.025	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.040%
50) Toluene-d8	8.647	98	351868	45.799	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.600%#
62) 4-Bromofluorobenzene	11.079	95	150192	48.835	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.680%
Target Compounds						
44) Trichloroethene	7.129	130	39085	18.864	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
 Data File : VX038847.D
 Acq On : 16 Nov 2023 18:58
 Operator : JC/MD
 Sample : 05241-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-4-20231102

Quant Time: Nov 16 23:06:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

