

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
 Data File : VX038848.D
 Acq On : 16 Nov 2023 19:22
 Operator : JC/MD
 Sample : 05357-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP01-20231107

Quant Time: Nov 16 23:06:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132685	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	275295	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	278862	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	141453	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99716	46.640	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.280%
35) Dibromofluoromethane	5.385	113	88088	44.304	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.600%
50) Toluene-d8	8.647	98	347169	45.637	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.280%#
62) 4-Bromofluorobenzene	11.079	95	148730	48.841	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.680%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1684	1.028	ug/l	94
12) 1,1-Dichloroethene	2.325	96	2743	1.659	ug/l #	85
27) cis-1,2-Dichloroethene	4.501	96	2628	1.236	ug/l	76
30) Chloroform	5.105	83	1748	0.527	ug/l #	76
38) Carbon Tetrachloride	5.684	117	3801	1.443	ug/l #	80
44) Trichloroethene	7.123	130	209991	102.359	ug/l	98
52) Toluene	8.720	92	1333	0.255	ug/l	94
55) 1,1,2-Trichloroethane	9.159	97	1532	0.702	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
Data File : VX038848.D
Acq On : 16 Nov 2023 19:22
Operator : JC/MD
Sample : 05357-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP01-20231107

Quant Time: Nov 16 23:06:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

