

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
 Data File : VX038851.D
 Acq On : 16 Nov 2023 20:31
 Operator : JC/MD
 Sample : 05403-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB02-20231107

Quant Time: Nov 16 23:07:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	130810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	266412	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268902	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98298	46.635	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.280%
35) Dibromofluoromethane	5.385	113	84900	44.125	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.240%
50) Toluene-d8	8.647	98	337396	45.831	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.660%#
62) 4-Bromofluorobenzene	11.079	95	144344	48.981	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.960%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	10520	9.376	ug/l	90
20) Methylene Chloride	2.788	84	3565	1.762	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
Data File : VX038851.D
Acq On : 16 Nov 2023 20:31
Operator : JC/MD
Sample : 05403-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB02-20231107

Quant Time: Nov 16 23:07:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

