

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
 Data File : VX038857.D
 Acq On : 16 Nov 2023 22:50
 Operator : JC/MD
 Sample : 05357-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT161S1-20231108

Quant Time: Nov 17 03:21:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	263552	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	272680	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135373	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96580	46.394	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.780%
35) Dibromofluoromethane	5.385	113	83555	43.897	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.800%
50) Toluene-d8	8.647	98	334953	45.993	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.980%#
62) 4-Bromofluorobenzene	11.079	95	144339	49.511	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.020%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	1366	0.848	ug/l #	76
24) 1,1-Dichloroethane	3.611	63	9773	2.955	ug/l	98
44) Trichloroethene	7.135	130	1925	0.980	ug/l	99
64) Tetrachloroethene	9.275	164	719	0.424	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
Data File : VX038857.D
Acq On : 16 Nov 2023 22:50
Operator : JC/MD
Sample : 05357-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT161S1-20231108

Quant Time: Nov 17 03:21:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

