

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
 Data File : VX038860.D
 Acq On : 16 Nov 2023 23:59
 Operator : JC/MD
 Sample : 05403-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20231107

Quant Time: Nov 17 03:22:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126328	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	262148	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	264296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131188	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96457	47.385	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.780%
35) Dibromofluoromethane	5.385	113	83736	44.228	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.460%
50) Toluene-d8	8.647	98	331463	45.758	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.520%#
62) 4-Bromofluorobenzene	11.079	95	140509	48.455	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.920%
Target Compounds						
52) Toluene	8.720	92	1463	0.294	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
Data File : VX038860.D
Acq On : 16 Nov 2023 23:59
Operator : JC/MD
Sample : 05403-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20231107

Quant Time: Nov 17 03:22:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

