

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
 Data File : VX038839.D
 Acq On : 16 Nov 2023 15:33
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX111623

Quant Time: Nov 16 23:01:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.583	0.601	-3.1	90	0.00
3 P	Chloromethane	0.761	0.709	6.8	90	0.00
4 C	Vinyl Chloride	0.734	0.703	4.2#	90	0.00
5 T	Bromomethane	0.303	0.310	-2.3	83	0.00
6 T	Chloroethane	0.419	0.403	3.8	89	0.00
7 T	Trichlorofluoromethane	0.931	0.893	4.1	90	0.00
8 T	Diethyl Ether	0.405	0.366	9.6	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.617	0.584	5.3	89	0.00
10 T	Methyl Iodide	0.751	0.733	2.4	89	0.00
11 T	Tert butyl alcohol	0.224	0.232	-3.6	99	0.00
12 CM	1,1-Dichloroethene	0.623	0.581	6.7#	88	0.00
13 T	Acrolein	0.119	0.105	11.8	89	0.00
14 T	Allyl chloride	1.248	1.182	5.3	90	0.00
15 T	Acrylonitrile	0.473	0.466	1.5	92	0.00
16 T	Acetone	0.429	0.406	5.4	90	0.00
17 T	Carbon Disulfide	1.930	1.721	10.8	89	0.00
18 T	Methyl Acetate	2.037	1.985	2.6	91	0.00
19 T	Methyl tert-butyl Ether	2.192	2.079	5.2	89	0.00
20 T	Methylene Chloride	0.773	0.687	11.1	90	0.00
21 T	trans-1,2-Dichloroethene	0.688	0.636	7.6	89	0.00
22 T	Diisopropyl ether	2.289	2.179	4.8	89	0.00
23 T	Vinyl Acetate	1.649	1.594	3.3	88	0.00
24 P	1,1-Dichloroethane	1.280	1.203	6.0	89	0.00
25 T	2-Butanone	0.696	0.669	3.9	91	0.00
26 T	2,2-Dichloropropane	1.110	1.019	8.2	87	-0.01
27 T	cis-1,2-Dichloroethene	0.802	0.742	7.5	89	0.00
28 T	Bromochloromethane	0.615	0.597	2.9	93	0.00
29 T	Tetrahydrofuran	0.442	0.439	0.7	91	0.00
30 C	Chloroform	1.250	1.161	7.1#	89	0.00
31 T	Cyclohexane	1.150	1.092	5.0	90	0.00
32 T	1,1,1-Trichloroethane	1.083	1.001	7.6	89	-0.01
33 S	1,2-Dichloroethane-d4	0.806	0.743	7.8	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.361	0.333	7.8	85	0.00
36 T	1,1-Dichloropropene	0.518	0.481	7.1	88	0.00
37 T	Ethyl Acetate	0.674	0.655	2.8	90	0.00
38 T	Carbon Tetrachloride	0.478	0.456	4.6	90	0.00
39 T	Methylcyclohexane	0.675	0.631	6.5	89	0.00
40 TM	Benzene	1.531	1.446	5.6	89	0.00
41 T	Methacrylonitrile	0.350	0.343	2.0	88	0.00
42 TM	1,2-Dichloroethane	0.507	0.482	4.9	89	0.00
43 T	Isopropyl Acetate	1.078	1.034	4.1	89	0.00
44 TM	Trichloroethene	0.373	0.351	5.9	88	0.00
45 C	1,2-Dichloropropane	0.412	0.395	4.1#	88	0.00
46 T	Dibromomethane	0.275	0.263	4.4	88	0.00
47 T	Bromodichloromethane	0.545	0.508	6.8	89	0.00
48 T	Methyl methacrylate	0.523	0.478	8.6	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
 Data File : VX038839.D
 Acq On : 16 Nov 2023 15:33
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX111623

Quant Time: Nov 16 23:01:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.011	-10.0	100	0.00
50 S	Toluene-d8	1.382	1.298	6.1	87	0.00
51 T	4-Methyl-2-Pentanone	0.689	0.674	2.2	90	0.00
52 CM	Toluene	0.950	0.890	6.3#	88	0.00
53 T	t-1,3-Dichloropropene	0.648	0.613	5.4	89	0.00
54 T	cis-1,3-Dichloropropene	0.669	0.641	4.2	88	0.00
55 T	1,1,2-Trichloroethane	0.396	0.368	7.1	87	0.00
56 T	Ethyl methacrylate	0.671	0.648	3.4	88	0.00
57 T	1,3-Dichloropropane	0.656	0.619	5.6	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.312	0.316	-1.3	111	0.00
59 T	2-Hexanone	0.552	0.550	0.4	91	0.00
60 T	Dibromochloromethane	0.403	0.383	5.0	89	0.00
61 T	1,2-Dibromoethane	0.416	0.392	5.8	88	0.00
62 S	4-Bromofluorobenzene	0.553	0.512	7.4	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.311	0.294	5.5	90	0.00
65 PM	Chlorobenzene	1.077	1.013	5.9	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.367	5.2	90	0.00
67 C	Ethyl Benzene	1.979	1.888	4.6#	89	0.00
68 T	m/p-Xylenes	0.749	0.719	4.0	90	0.00
69 T	o-Xylene	0.741	0.702	5.3	89	0.00
70 T	Styrene	1.251	1.191	4.8	89	0.00
71 P	Bromoform	0.323	0.317	1.9	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.896	3.672	5.7	89	0.00
74 T	N-amyl acetate	2.059	2.029	1.5	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.521	1.433	5.8	89	0.00
76 T	1,2,3-Trichloropropane	1.247	1.251	-0.3	93	0.00
77 T	Bromobenzene	0.875	0.819	6.4	89	0.00
78 T	n-propylbenzene	4.795	4.572	4.7	89	0.00
79 T	2-Chlorotoluene	2.846	2.667	6.3	89	0.00
80 T	1,3,5-Trimethylbenzene	3.276	3.162	3.5	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.562	0.546	2.8	88	0.00
82 T	4-Chlorotoluene	3.281	3.127	4.7	89	0.00
83 T	tert-Butylbenzene	3.226	3.045	5.6	89	0.00
84 T	1,2,4-Trimethylbenzene	3.315	3.159	4.7	89	0.00
85 T	sec-Butylbenzene	4.220	4.029	4.5	89	0.00
86 T	p-Isopropyltoluene	3.387	3.232	4.6	90	0.00
87 T	1,3-Dichlorobenzene	1.714	1.621	5.4	89	0.00
88 T	1,4-Dichlorobenzene	1.754	1.623	7.5	88	0.00
89 T	n-Butylbenzene	3.418	3.289	3.8	89	0.00
90 T	Hexachloroethane	0.676	0.656	3.0	90	0.00
91 T	1,2-Dichlorobenzene	1.626	1.567	3.6	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.373	0.360	3.5	88	0.00
93 T	1,2,4-Trichlorobenzene	1.089	1.038	4.7	89	0.00
94 T	Hexachlorobutadiene	0.386	0.376	2.6	90	0.00
95 T	Naphthalene	4.058	3.835	5.5	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX111623\
Data File : VX038839.D
Acq On : 16 Nov 2023 15:33
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX111623

Quant Time: Nov 16 23:01:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.033	0.997	3.5	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6